



## Basic UDI-DI & UDI-DI attributes

### Basic UDI-DI set of data in UDI database

Principle: Each UDI-DI inherits the attributes of its linked Basic UDI-DI and devices DI

#### Basic UDI-DI

Applicable legislation (IVDR) (\*)

- 2. Basic UDI-DI value (\*)
- 2b Basic UDI-DI Issuing entity (\*);
- 6. Manufacturer SRN (\*)
- 5. Name and address of manufacturer
- 7. Name and address and SRN of AR
- 9. Risk class (\*)
  - A.2.14 Intended for self-testing (Y/N) (\*)
  - A.2.14 Intended for near-patient-testing (Y/N) (\*)
- Companion diagnostics (Y/N) (\*)
- Instrument(Y/N) (\*)
- Reagent(Y/N) (\*)
- Professional testing (Y/N) (\*)
- 11. A. Name and/or, if applicable, device model that identifies the device(s) with this BASIC UDI-DI in the technical documentation and/or certificate or declaration of conformity  
*(Name and/or model shall be provided)*

#### UDI-DIs

- 0. UDI-DI value (\*)
- 0b. UDI-DI Issuing entity (\*)
- Secondary DI (value and issuing entity)
- 11.B. Reference, Article or Catalogue number (\*)
- Device with Direct marking (Y/N) (\*)
- Direct marking UDI-DI value (\*)
- Direct marking UDI-DI issuing entity (\*)
- 1. Quantity of device(s) (\*)
- 3. Type of UDI-PI (\*)
- 4. Unit of use UDI-DI (\*)
- 13. Storage/handling conditions
- 10-14. Name(s)/Trade name(s) (including languages)
- 12. Additional product description
- 19. URL for additional information
- 15. Labelled as single use (YN) (\*)
- 16. Maximum number of reuse (\*)
- 17. Device labelled sterile (Y/N) (\*)
- 18. Need for sterilisation (Y/N) (\*)
- 20. Critical warnings or contra-indications
- 8. Medical device nomenclature (CND) code (1)
- 21. Status
- 27 (A.2.10). In the case of devices designed and manufactured by another legal or natural person as referred in Article 10(14), the name, address and contact details of that natural/legal person

#### UDI-DIs (container package DI)

- 0. UDI-DI value (\*)
- 0b. Issuing entity (\*)
- 1. Quantity per package (\*)
- 21. Status

(1) Nomenclature decision:

[https://ec.europa.eu/doc\\_sroom/documents/34264](https://ec.europa.eu/doc_sroom/documents/34264)

(\*) may not be changed

- Mandatory
- Mandatory if applicable
- Optional

## Other Device Data attributes

### Basic UDI-DI

- **A.2.2 Certificate IDs (with NB, type .. Link);**
- **A.2.11 SSP;**
- **A.2.9 Performance study IDs (..link);**
- **A.2.5 Presence of Human tissues/Cells (Y/N) (\*);**
- **A.2.6 Presence of Animal tissues/Cells (Y/N) (\*);**
- **A.2.7 Presence of Substances/cells of microbial origin (Y/N) (\*);**
- **Kit (Y/N) (\*);**




 Provided by NB or for certificate ID under Art 26(2) provided by manufacturer and confirmed by NB

**Version April 2019**

### UDI-DIs

- **A.2.13 New Device (Y/N) (\*);**
- **A.2.3 Member State of the Placing on the EU Market of the Device (\*);**
- **A.2.4 Member State(s) where the Device is made available in the Country;**

(\*) may not be changed

-  **Mandatory**
-  **Mandatory if applicable**
-  **Optional**